



中国认可
国际互认
检测
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

检测报告(Testing Report)	防伪查询网址(Security website): www.fcl-sz.org.cn 防伪码(Security code): 3680011725
报告编号(Report No): ZFLJ2641963	Page 1 of 4



Applicant Information (客户信息)

Applicant Name(委托单位) : Yiwu Lanhu leather Co., LTD(义乌市蓝狐皮具有限公司)
Applicant Address(委托单位地址) : No.138, Dongyuan Road, Qingkou Industrial Zone, Jiangdong Street, Yiwu city, Zhejiang Province, China
(浙江省义乌市江东街道青口工业区东苑路 138 号)
Manufacturer(生产单位) : Yiwu Lanhu leather Co., LTD(义乌市蓝狐皮具有限公司)

Sample Information (样品信息)

Sample Description(样品描述) : KN95 Mask(KN95 口罩)
Style No.(款号) : juye991
Sample Quantity(样品数量) : 50 pieces(50 个)
- Sample Receiving date (收到样品的日期) : 2020-08-12
- Report date (报告日期) : 2020-08-14
- The original sample is stucked on the last paper. (送检样品原样粘贴在本报告的末页)

Test Performed (检验标准)

Judgement according to(评定依据):
Refer to(参照)GB 2626-2006 Respiratory protective equipment— Non-powered air-purifying particle respirator
呼吸防护用品 自吸过滤式防颗粒物呼吸器
- Selected test(s) as requested by applicants. (本实验室根据客户要求完成以下检测内容)

Pronounce (声明)

The results shown in this report refer only to sample(s) tested unless otherwise stated.All the items are performed in the standard conditions, except the noted cases.(除非特别说明,测试结果只对来样负责。所有测试项目均在标准规定的环境下进行,有注明除外。)
Except for the requirement of the client, the test results and the conformity judgement of this report do not take the uncertainty of the test results into account.(除非客户要求,本报告检测结果及符合性判定不考虑测量结果的不确定度。)

Signed for and on behalf of

CNTAC Testing Service Co.,Ltd.(Foshan)

Approved by
(批准)

张荣



检测报告(Testing Report)

防伪查询网址(Security website): www.fcl-sz.org.cn
防伪码(Security code): 3680011725

报告编号(Report No): ZFLJ2641963

Page 2 of 4

--- Test Result (测试结果) ---

Test Result(测试结果)

Requirements(评定条件)

Judgement(判定)

1. Filtration Efficiency(Salt medium) (过滤效率(盐性介质)) GB 2626-2006 Section 6.3(条款 6.3)

Air flow: 85L/min, Aerosol: NaCl(气体流量 85L/min, 气溶胶颗粒: NaCl)

Unit(单位): <%>

Before pretreatment(未经处理)

(KN95) ≥ 95.0

Pass(合格)

Sample 1#(样品 1) 99.3

sample 2#(样品 2) 99.6

sample 3#(样品 3) 99.2

sample 4#(样品 4) 99.7

sample 5#(样品 5) 99.6

sample 6#(样品 6) 99.3

sample 7#(样品 7) 99.4

sample 8#(样品 8) 99.5

sample 9#(样品 9) 99.7

sample 10#(样品 10) 99.8

Minimum(最小值) 99.2

2. Exhalation Resistance(呼气阻力) GB 2626-2006 Section 6.6(条款 6.6)

Air flow: 85 L/min(通气量 85 L/min)

Unit(单位): <Pa>

Before pretreatment(未经处理)

 ≤ 250

Pass(合格)

sample 1#(样品 1) 72.4

sample 2#(样品 2) 85.7

Maximum(最大值) 85.7

3. Inhalation Resistance(吸气阻力) GB 2626-2006 Section 6.5(条款 6.5)

Air flow: 85 L/min(通气量 85 L/min)

Unit(单位): <Pa>

Before pretreatment(未经处理)

 ≤ 350

Pass(合格)

sample 1#(样品 1) 92.8

sample 2#(样品 2) 100.1

Maximum(最大值) 100.1



检测报告(Testing Report)	防伪查询网址(Security website): www.fcl-sz.org.cn 防伪码(Security code): 3680011725	
报告编号(Report No): ZFLJ2641963	Page 3 of 4	

Test Result(测试结果)	Requirements(评定条件)	Judgement(判定)
4. Appearance Inspection(外观检查) <u>GB 2626-2006 Section 6.1~6.2(条款 6.1~6.2)</u>		
Accordance(符合)	The surface should not be damaged, deformation and other obvious defects. (样品表面不应破损、变形和有明显的其他缺陷。)	Pass(合格)
Accordance(符合)	Component materials and structures shall be able to withstand normal operating conditions and possible temperature, humidity and mechanical shocks. (部件材料和结构应能耐受正常使用条件及可能遇到的温度、湿度和机械冲击。)	
Accordance(符合)	Headband should be adjustable. (头带应可调)	

Remark(备注): The client requires that only before pretreatment results be issued, And according to GB 2626-2006 standard. (客户要求只出具未处理结果, 并按 GB 2626-2006 标准判定。)



检测报告(Testing Report)


防伪查询网址(Security website): www.fcl-sz.org.cn
防伪码(Security code): 3680011725



报告编号(Report No): ZFLJ2641963

Page 4 of 4

Original (送检样品原样)



===== End of Report (检测报告结束) =====

